



RELEASE NOTES

Product: **FW for MI 2883_HW2.0_3.0**

Version: **1.0.2364**

Release date: 22.10.2019

DESCRIPTION OF THE RELEASED PRODUCT:

FW for MI 2883 Energy Master

DESCRIPTION OF CHANGES, DEPENDENCES AND HISTORY:

MI 2883				PC SW	
#	FW version	Release date	Description of changes:	HW version	PowerView3
6	1.0.2364	22.10.2019	<ul style="list-style-type: none"> • New dealer integration 	HW2_3	V 3.0.0.4268
5	1.0.2351	10.09.2019	<ul style="list-style-type: none"> • Added "Calibration date" info under "INSTRUMENT INFO" screen • Bug fixing 	HW2_3	V 3.0.0.4199
4	1.0.2340	07.12.2018	<ul style="list-style-type: none"> • Corrected power presentation for 3W connection • Support for A 1636 current clamps • Improved Smart clamp recognition procedure 	HW2_3	V 3.0.0.3797
3	1.0.2331	06.09.2018	<ul style="list-style-type: none"> • Fully support for daily and weekly IEEE 519 reporting • Added screenshot copy to picture file (.BMP), • Added automatic factory reset after firmware upgrade, • Added new current clamps, (A 1636 (AC/DC current clamps), A 1501/1502/1503 smart flex clamps, A 1588 – smart iron leakage clamps), • Added selection Flat/Folder data structure, • Added new dealers. 	HW2_3	V 3.0.0.3797
2	1.0.2221	20.12.2016	<ul style="list-style-type: none"> • Modified connection check tool (extended limits), • Added auto-ranging and auto setting of current range, for current clamps, • Added 2 additional measuring methods for Power measurement (Classic, <i>Vector and Arithmetic</i>), • Added profile setting (Limited) in General recorder, enabling support for 7days recording with 1sec interval, • Added VFD measurement, • Added option to name the instrument, • Added transformer K-factor measurement, • Added Thai language. 	HW2_3	V 3.0.0.2461

1	1.0.2121	19.07.2016	Bug fixes	HW2_3	V 3.0.0.2363
---	----------	------------	-----------	-------	--------------

INSTALLATION PROCEDURE:

The instrument can be upgraded from a PC via the RS232 or USB communication port. This enables to keep the instrument up to date even if the standards or regulations change. The firmware upgrade requires internet access and can be carried out from the **PowerView3** software with a help of special upgrading software – **FlashMe** which will guide you through the upgrading procedure. For more information refer to PowerView Help file.

COMPANY DATA:

METREL d.d.
Measuring and Regulation Equipment Manufacturer

Ljubljanska 77
SI-1354 Horjul

Tel: + 386 (0)1 75 58 200
Fax: + 386 (0)1 75 49 226
E-mail: metrel@metrel.si
<http://www.metrel.si>